

Common Lab Inspection Findings

Training

- PIs and Lab Personnel must complete all training required by Research Safety.

Labeling

- All chemical containers and storage areas must be properly labeled.

Check Safety Equipment

- The Safety Shower must be tested monthly, and the Eyewash weekly.
- Fire Extinguisher visual inspections must be done monthly.

Chemical Segregation

- Properly segregate incompatible chemicals. A storage guide can be found on the RS website.



<http://www.clemson.edu/research/safety>

Remember

Update Lab
Door Signs

Read & Sign
the CHP

Gas
Cylinders:
Chains not
Straps

Contact RS
before use
of
respirators/
dust masks